

# **Anonymous Anonymous**

On page 40 of 47, Table 1 please include other methods for BOD and Soluble BOD. Include the ability to use newer technology such as optical D.O. meters. These methods would include:  
In-Situ 1003-8-2009 for BOD.

ASTM D88-09C for Dissolved Oxygen.

It is only changing a probe to a more efficient kind. It currently takes the accreditation division of Ecology way too long to approve any method changes. Please just include EPA approved alternative methods now to save time and resources moving forward with all new permits.