



New Mexico Environment Department (NMED)

New Mexico Clean Transportation Fuel Program (CTFP) | Stakeholder Feedback

This comment is intended to recommend the use of the carbon-14 testing method to determine the share of biogenic carbon content of feedstocks, fuels and emissions under New Mexico's Clean Transportation Fuel Program. Biogenic content measurements following methods such as ASTM D6866 Method B currently provide critical value to existing state, federal and international clean fuel standard programs.

Included here you will find:

Recommendations for New Mexico's Clean Transportation Fuel Program	1
What is Biogenic Testing (Carbon-14)?	8
ASTM D6866 Method B - The Most Reliable Method	9
About Beta Analytic	10
ISO/IEC 17025:2017 laboratory	11
Required tracer-free facility for Carbon-14	11
References	11

Recommendations for New Mexico's Clean Transportation Fuel Program

Our recommendation is that New Mexico's Clean Transportation Fuel Program should include direct biogenic content testing (carbon-14) requirements following the ASTM D6866 Method B standard for any fuels or feedstocks seeking recognition of renewable (biogenic) content. Routine direct biogenic testing requirements are the only reliable method of incentivizing the use of biomass-derived content while guaranteeing compliance, and currently play a critical role in prominent similar programs. This comment follows up on our initial feedback submitted in January.

Introduce Routine Biogenic Testing Requirements

Our first recommendation is that NMED should introduce routine biogenic testing requirements in line with those in place for the US RFS, California's LCFS, Oregon's CFP and Washington's CFS. Introducing routine testing would also be in line with best practices established by Canada's CFR and the EU's RED.



ISO/IEC 17025:2017-Accredited Testing Laboratory

Routine direct test results are currently used to verify biogenic content under the US EPA's [Renewable Fuel Standard \(RFS\)](#), California's [Low Carbon Fuel Standard \(LCFS\)](#), Oregon's [Clean Fuels Program](#), Washington's [Clean Fuel Standard \(CFS\)](#), Canada's [Clean Fuel Regulations \(CFR\)](#) and the EU's [Renewable Energy Directive \(RED\)](#). All of these programs, except the EU RED specifically require the carbon-14 standard ASTM D6866, while the EU RED accepts ASTM D6866 or its European equivalents. ASTM D6866 is also required for prominent third-party verification programs, most notably the Roundtable on Sustainable Biomaterials (RSB).¹ Testing requirements allow clean fuel programs to exclusively incentivize the renewable portion of fuels. This is especially important given the recent history of attempted fraud in existing transportation fuel decarbonization programs.

New Mexico's Clean Transportation Fuel Program should require direct biogenic testing for any fuels produced from co-processing, municipal solid waste (MSW) biogas & renewable natural gas (RNG) and any other fuels for which the final biogenic content is unknown. Current requirements of routine direct testing following ASTM D6866 under similar prominent programs include (please see specific rules hyperlinked):

- The US RFS currently [requires](#) routine direct testing following ASTM D6866 for fuels produced from co-processing, municipal solid waste (MSW), [biogas and renewable natural gas \(RNG\)](#).²
- California's LCFS [requires](#) routine direct testing for fuels produced from co-processing and recommends for fuels produced from MSW.³
- Oregon's CFP [requires](#) routine direct testing following the protocols of the US RFS third-party engineering reviews.⁴
- Washington's CFS [requires](#) routine direct testing following the protocols of the US RFS third-party engineering reviews.⁵
- Canada's CFR [requires](#) routine direct testing for any fuels produced from co-processing and their co-products.⁶
- British Columbia's LCFS [requires](#) monthly testing for any fuels produced from co-processing and quarterly testing for their co-products, as well as to verify biogenic feedstocks.⁷
- The EU's RED [requires](#) routine direct testing for any fuels produced from co-processing or biogas and renewable natural gas (RNG).⁸

Always Require Calculations to be Verified by Direct Testing

¹ 2023. "RSB Standard for Advanced Fuels." *Roundtable on Sustainable Biomaterials (RSB)*

² 2023. "40 CFR Parts 80 and 1090– Renewable Fuel Standard (RFS) Program: Standards for 2023–2025 and Other Changes." *EPA*

³ 2020. "Reporting Co-Processing and Renewable Gasoline Emissions Under MRR." *California Air Resources Board*

⁴ 2023. "Oregon Clean Fuels Program." *Oregon Department of Environmental Quality*

⁵ 2022. "Chapter 173-424 WAC: Clean Fuels Program Rule." *Washington State Legislature*

⁶ 2022. "Clean Fuel Regulations: Quantification Method for Co-Processing in Refineries." *Environment and Climate Change Canada*

⁷ 2025. "Low Carbon Fuel Regulation: Co-Processing Methodology" *British Columbia Ministry of Energy and Climate Solutions*

⁸ 2023. "Renewable energy- method for calculating the share of renewables in the case of co-processing." *European Commission*



ISO/IEC 17025:2017-Accredited Testing Laboratory

Beta believes it is not in the best interest of New Mexico's CTFP to allow any mass balance calculations to be used for reporting biogenic content under this program. If NMED does allow any mass balance calculations, it is critically important to require these calculations to be verified by routine direct testing. We stress the importance of reviewing other programs' experiences with these calculation-based approaches to understand the risk they would introduce to the program.

Producers and industry lobbying groups favour calculation-based approaches such as mass balance because they enable facilities to make claims solely based on material inputs in production. These calculations allow producers to assume that all of their biomass inputs end up in their facilities' outputs, despite it being well understood in the industry that the input of renewable feedstocks is not the same as the output because performance varies and renewable feedstocks don't produce the same quantity of material as their fossil counterparts.⁹ By basing their calculations solely on production inputs rather than outputs these methods systematically over-report the renewable share of fuels.

Calculation-based approaches also use a system of free allocation, meaning they do not have to guarantee that there is any renewable content in a given fuel. Producers prefer this because if 10% of their feedstocks are biogenic they can claim that 10% of their products are biogenic, even if that's not the case because biobased can go in different amounts to different products in the co-process. Even further, book and claim also allows them to claim that 10% of their products are 100% biogenic and the rest are 0%, even if all of the products should be 10% biogenic based on calculations (and would likely C14 test below that).¹⁰

These calculations' reliance on free allocation creates the potential for double counting of renewable content, leaving low carbon fuel programs susceptible to a high risk of greenwashing and fraud. For example, this threat is highlighted by the recent mass balance fraud challenges faced by the ISCC regarding fraudulent biodiesel submissions from China which "caused a dramatic fall in biodiesel prices in European markets" in July 2023.¹¹ In response to this situation, the EU quickly updated the RED's co-processing rules to uniformly require direct testing, including verifying the calculations of producers choosing to use calculation-based approaches.¹²

The importance of limiting the role of mass balance for reporting the biogenic content of fuels is articulated very well by a [recent judgement](#) of the Advocate General of the EU Court of Justice (CJEU) on the roles of mass balance and C-14 for reporting biogenic content in co-processing. The official opinion found that mass balance calculations are not intended to quantify the share of biogenic contained in a

⁹ 2006. "Determining the modern carbon content of biobased products using radiocarbon analysis." *Bioresource Technology*, 97(16), 2084-2090.

¹⁰ 2024. "The Mass Balance Approach." *International Sustainability & Carbon Certification*

¹¹ 2023. "ISCC Press Release July 27, 2023." *International Sustainability & Carbon Certification*

¹² 2023. "Renewable energy- method for calculating the share of renewables in the case of co-processing." *European Commission*



ISO/IEC 17025:2017-Accredited Testing Laboratory

biofuel produced by co-processing.¹³ This judgment was issued in response to a case brought by BP France against the French government regarding a tax incentive requiring C-14 testing to verify claims of renewable content. BP is also notably a board member of the ISCC.¹⁴

Recently, in the US, issues with mass balance in the recycling industry have received increasing attention. A [ProPublica investigation published in June 2024](#) found that products advertised as 30% recycled through mass balance often contained less than 1% recycled content.¹⁵ Similar concerns were shown by the US EPA as early as 2023, which described the mass-balance methodology as deceptive and advised against promoting it. In August 2024, the US Environmental Protection Agency (EPA) launched a federal action against the mass-balance methodology used in the recycling sector.

In September 2024, California Attorney General Rob Bonta filed [a lawsuit against ExxonMobil](#) claiming that the oil major “deceptively” promoted chemical recycling as a solution to the plastic crisis, citing their use of mass balance calculations such as ISCC Plus.¹⁶ That lawsuit directly challenges the standard’s use of ISCC’s free allocation method as a system designed to enable greenwashing.¹⁷ The New York Times also recently [published a relevant article](#) on the challenges that mass balance presents to the recycling industry, which aligns with the challenges experienced in the renewable products industry.¹⁸

Beta specifically encourages NMED to reconsider its proposed use of mass balance under 20.2.92.504 (J) for quarterly reporting of comingled storage transactions. This would be an ineffective approach to monitoring the use of biogenic content from comingled fuels under the program, based on the issues with the mass balance system discussed above. For the same reasons that mass balance cannot be used to attribute the biogenic content of co-processed fuels, it is not an appropriate methodology for attributing the biogenic content of comingled fuels. Once a renewable fuel is comingled in a common storage tank, only direct biogenic testing using C-14 analysis can accurately determine the renewable share of the fuels withdrawn. The same goes for biomethane distributed through a comingled pipeline as discussed under 20.2.92.504 (J)(2).

It is in the best interest of New Mexico’s decarbonization goals not to allow any producers to report their biogenic content using mass balance calculations. However, if mass balance is used at all it is critical that these calculations be routinely verified by direct testing. The advantage of the updated RED protocol is that producers can choose to use calculations internally, while the program still ensures the information

¹³ 2024. “Opinion of Advocate General Campos Sánchez-Bordona Delivered on 11 January 2024: Case C-624/22.” *Court of Justice of the European Union*

¹⁴ 2024. “Board Members of the ISCC Association.” *International Sustainability & Carbon Certification*

¹⁵ 2024. “Biden EPA Rejects Plastics Industry’s Fuzzy Math That Misleads Customers About Recycled Content.” *ProPublica*

¹⁶ 2024. “The People of the State of California v. Exxon Mobil Corporation.” *Superior Court of the State of California*

¹⁷ 2024. “ExxonMobil Accused of “Deceptively” Promoting Chemical Recycling as a Solution for the Plastics Crisis.” *ProPublica*

¹⁸ 2024. “Is Your Water Bottle Really Made From Recycled Plastic?” *The New York Times*



reported is accurate through direct Carbon-14 analysis. This is the only way to mitigate the risk to the program introduced by these calculations.

Require Routine Testing for Co-Processing in Refineries

The proposed requirements for co-processing refineries in the public review draft are insufficient to regulate co-processed fuels in this program. Under subsection D of 20.2.92.205 NMAC, routine direct testing following ASTM D6866 Method B should be the only accepted method for reporting the biogenic proportions of co-processed fuels and co-products and should be used to determine their carbon intensity. The language included in this section reflects the same sections of the laws establishing Oregon's CFR and Washington's CFS, both of which require Carbon-14 test results to establish the fraction of co-processed products that can be claimed as renewable in quarterly reporting, and to fill out the fuel pathway application and annual report. NMED should use this section to implement those same requirements to ensure accurate reporting of biogenic content for any co-processing under this program.

As discussed above, refineries conducting co-processing are required to verify the renewable portion of their fuels under the US RFS, California's LCFS, Oregon's CFP, Washington's CFS, Canada's CFR and the EU RED. We re-emphasize the importance of the EU RED as a relevant example which allowed co-processed fuels to be submitted exclusively using calculations and was forced to quickly adopt C-14 testing requirements after discovering a massive case of fraudulent fuels last year.

In addition to the regulations for other programs linked above in the, "Introduce Routine Biogenic Testing Requirements," section, we also urge NMED to review the following studies on co-processing conducted by the ASTM D02 Committee on petroleum products, liquid fuels and lubricants. We specifically recommend reviewing RR:D02-2052, which compares the results of ¹⁴C and mass balance in co-processing facilities.¹⁹ The table below shows an example of that study's key findings.

¹⁹ 2023. "RR:D02-2052." *ASTM International*



ISO/IEC 17025:2017-Accredited Testing Laboratory

Standard	Report #					
D1655	RRD02-2052					
Findings						
Pre-Fractionation Blend Analysis						
	Component, vol %	Blend 1 Pre	Blend 2 Pre	Blend 3Pre		
	Petrochemical Stream	60	80	80	<i>This Pre-Fractionation verification of biogenic content of the blends demonstrates the accuracy of ASTM D6866</i>	
	Feed 1	40				
	Feed 2		20			
	Feed 3			20		
	D6866 Method B	40	20	20		
Biogenic %, per ASTM D6866 Method B	Blend 1 Pre	Blend 1 Post	Blend 2 Pre	Blend 2 Post	Blend 3 Pre	Blend 3 Post
	40	17	20	7	20	5
Standard	Report #					
D1655	RRD02-1886					
Findings						
Biomass Input 5%, yield in final product by ASTM D6866 2.1%						

The black font in the table shows the expected values of biogenic content based on mass balance calculations, while the red shows the actual values reported by direct testing. The study shows that mass balance consistently over-estimated the biogenic content which ended up in co-processed fuels because biomass does not behave the same as fossil feedstocks, and not all of the biomass inputs necessarily end up in the same output.²⁰

We recommend further reviewing RR:D02-2052, as well as the rest of this collection of technical reports which includes RR:D02-1886, RR:D02-1929, RR:D02-2052, RR:D02-1739, RR:D02-1810, RR:D02-1776, RR:D02-1884, RR:D02-1828, and RR:D02-2039. Several of these studies specifically compare the results of ¹⁴C and mass balance in co-processing facilities in the context of sustainable aviation fuel production in particular. These studies found that mass balance calculations are consistently unable to estimate the renewable portion of co-processed fuels and should not be relied on as the sole method of verification for clean fuel programs. All of these technical reports are available from ASTM upon request.

Require Testing for Any Biomethane Book-and-Claim

As the biomethane industry accelerates in jurisdictions with clean or low-carbon fuel programs, properly regulating the industry in this early stage is key to its future success. Recent developments in the US RFS and EU RED have demonstrated best practices for regulating biogas, biomethane and renewable natural gas (RNG) based on these programs' early experiences with these fuels.

Under Section 20.2.92.504 (4) on page 46 of this draft and 20.2.92.506 (11) on page 48, biogenic test results following ASTM D6866 should be required evidence of the "environmental attributes," of

²⁰ 2023. "RR:D02-2052." ASTM International



ISO/IEC 17025:2017-Accredited Testing Laboratory

biomethane necessary to use a book-and-claim methodology. This would also align with the requirement of 20.2.92.506 (11)(d) to provide documentation of claims made under the federal RFS for the same biomethane, because routine ASTM D6866 test results are already a federal requirement.

The only way to reliably differentiate biogenic biomethane from fossil fuel methane is to require mandatory routine test results following ASTM D6866 Method B for any entities seeking recognition of emission reductions from the use of biomethane. Since biomethane and fossil fuel methane are chemically identical molecules, the only way to differentiate the two is to perform carbon-14 testing of the fuels or the emissions after combustion to assess what percentage of the mixture was biogenic.

The EU introduced biogenic testing requirements for fuels produced from biogas in a June 2023 update to the EU Renewable Energy Directive (RED) titled, [“Renewable energy- method for calculating the share of renewables in the case of co-processing.”](#)²¹ This update was specifically issued in response to the discovery of a major case of fraud within the RED program stemming from biodiesel submissions from China which were approved by mass balance calculations.²² The EU investigation into this issue is still ongoing, and the full extent of the damage is not yet known, but this was a significant setback for the program and quickly plummeted biodiesel prices in the EU. The EU tied biogas and RNG into the update in order to address these concerns for any fuels containing a mixture of biogenic and fossil content.

The advantage of this framework is that the EU was able to continue to accept calculation-based methodologies like mass and energy balance by requiring routine direct biogenic testing to validate the data. However, calculation-based approaches are much more common for co-processing, where all inputs and outputs are concentrated in a single facility, as opposed to biomethane and RNG which are often produced, upgraded and blended at multiple facilities.

The US introduced biogenic testing requirements for fuels produced from biogas in the 2023 Set Rule update to the US Renewable Fuel Standard (RFS), in a section called the [Biogas Regulatory Reform Rule](#).²³ This update requires routine biogenic testing for any biogas or RNG fuels seeking to generate RINs under the RFS. Starting on July 1st, 2024 for new facilities and January 1st, 2025 for existing facilities, fuels produced from biogas will need to submit biogenic test results of the biogas at the point of production from the digester/landfill, at the point of upgrading, and after upgrading prior to pipeline injection.

Beta encourages NMED to consider following a similar approach to enable a book-and-claim system for biomethane using routine direct testing. The US RFS model of testing at the point of production, at the point of blending with non-renewable components and at the point of injection into a pipeline, provides

²¹ 2023. “Renewable energy- method for calculating the share of renewables in the case of co-processing.” *European Commission*

²² 2023. “ISCC Press Release July 27, 2023.” *International Sustainability & Carbon Certification*

²³ 2023. “40 CFR Parts 80 and 1090– Renewable Fuel Standard (RFS) Program: Standards for 2023–2025 and Other Changes.” *EPA*



a comprehensive chain of custody for the renewable content in these fuels, making it possible to report and trade only real biogenic content introduced to the grid. Similarly, the EU RED model demonstrates that tying calculation-based accounting approaches to routine direct testing is the most secure way to access the benefits of a book-and-claim system without exposing the program to undue risk.

Conclusion

Establishing this Clean Transportation Fuel Program was a critical first step in New Mexico's decarbonization journey. By implementing best practices for verification established by similar state, federal and international fuel decarbonization programs New Mexico's Environment Department can best prepare this program to successfully achieve and measure its goals. Routine direct testing following ASTM D6866 Method B is the most effective way to incentivize and validate biogenic content under this program.

What is Biogenic Testing (Carbon-14)?

Carbon-14 analysis is a reliable method used to distinguish the percentage of biobased carbon content in a given material. The radioactive isotope carbon-14 is present in all living organisms and recently expired material, whereas any fossil-based material that is more than 50,000 years old does not contain any carbon-14 content. Since Carbon-14 is radioactive, the amount of carbon-14 present in a given sample begins to gradually decay after the death of an organism until there is no carbon-14 left. Therefore, a radiocarbon dating laboratory can use carbon-14 analysis to quantify the carbon-14 content present in a sample, determining whether the sample is biomass-based, fossil fuel-derived, or a combination.

The analysis is based on standards such as ASTM D6866 and its international equivalents developed for specific end uses, such as ISO 13833. ASTM D6866 is an international standard developed for measuring the biobased carbon content of solid, liquid, and gaseous samples using radiocarbon dating.²⁴ There are also many international standards based on the specific use of direct Carbon-14 testing, such as ISO 13833, which is an international standard developed for measuring the biogenic carbon content of stationary sources emissions.²⁵

Carbon-14 analysis yields a result reported as % biobased carbon content. If the result is 100% biobased carbon, this indicates that the sample tested is completely sourced from biomass material such as plant or animal byproducts. A result of 0% biobased carbon means a sample is only fossil fuel-derived. A sample that is a mix of both biomass sources and fossil fuel sources will yield a result that ranges between 0% and 100% biobased carbon content. Carbon-14 testing has been incorporated into several

²⁴ 2021. "Standard Test Methods for Determining the Biobased Content of Solid, Liquid, and Gaseous Samples Using Radiocarbon Analysis." *ASTM International (D6866-21)*

²⁵ 2013. "ISO 13833:2013 Stationary source emissions: Determination of the ratio of biomass (biogenic) and fossil-derived carbon dioxide." *International Organization for Standardization*



regulations as the recommended or required method to quantify the biobased content of a given material.

ASTM D6866 Method B - The Most Reliable Method

Carbon-14 is a very well-established method which has been in use by many industries (including the fossil fuel industry) and academic researchers for several decades.

Carbon-14 measurements done by commercial third party testing is robust, consistent, and with quantifiable accuracy/precision of the carbon-14 amount under **ASTM D6866 method B**. The EN 16785 is the only standard that allows a variant of the Mass Balance (MB) method of ‘carbon counting’ under EN 16785-2. The EN 16785-1 requires that the biocarbon fraction be determined by the carbon-14 method. However, when incorporating this EN 16785 method, certification schemes like the “Single European Bio-based Content Certification” **only** allow the use of EN 16785-1 due to its reliability and the value of a third-party certification. <http://www.biobasedcontent.eu/en/about-us/>

In ASTM D6866 method B, the carbon-14 result is provided as a single numerical result of carbon-14 activity, with graphical representation that is easily understood by regulators, policy makers, corporate officers, and more importantly, the public. The overwhelming advantage of carbon-14 is that it is an independent and standardized laboratory measurement of any carbon containing substance that produces highly accurate and precise values. In that regard, it can stand alone as a quantitative indicator of the presence of biobased vs. petroleum feedstocks. When carbon-14 test results are challenged, samples can be rapidly remeasured to verify the original reported values (unlike mass balance).

The quantification of the biobased content of a given product can be as low as 0.1% to 0.5% (1 relative standard deviation – RSD) based on Instrumental error for Method B (AMS). This error is exclusive of indeterminate sources of error in the origin of the biobased content, and manufacturing processes. As such a total error of +/-3% (absolute) has been assigned to the reported Biobased Content to account for determinate and indeterminate factors.²⁶

It is also important that the program should always require ASTM D6866 Method B, rather than allow Method C for any use. Where ASTM D6866 Method B uses the AMS Instrument to measure ¹⁴C, Method C uses Liquid Scintillation Counting (LSC). In Method B, the AMS Instrument directly measures the ¹⁴C isotopes. However, in Method C, scintillation molecules indirectly absorb the beta molecules that release with the decay of ¹⁴C and convert the energy into photons which are measured proportionally to the amount of ¹⁴C in the sample. Since Method B directly measures the ¹⁴C isotopes and Method C measures

²⁶2021. Standard Test Methods for Determining the Biobased Content of Solid, Liquid, and Gaseous Samples Using Radiocarbon Analysis. *ASTM International (D6866-21)*. pp 1-19. doi: 10.1520/D6866-21.



them indirectly, Method B is significantly more precise and should be prioritized in regulations.²⁷ LSC measurements, like those used in Method C, are commonly used as an internal testing tool when samples are limited and accuracy does not need to be extremely high.

About Beta Analytic

Beta Analytic was among the originators of the use of Accelerator Mass Spectrometry (AMS) for the ASTM D6866 biobased / biogenic testing standard using Carbon-14 to distinguish renewable carbon sources from petroleum sources. Beta began testing renewable content in 2003 at the request of United States Department of Agriculture (USDA) representatives who were interested in Beta's Carbon-14 capabilities for their BioPreferred[®] Program (www.biopreferred.gov). At their request, Beta joined ASTM under subcommittee D20.96. Beta's previous president, Darden Hood, was positioned as a technical contact for the USDA and within 3 months completed the ASTM D6866-04 standard. The Carbon-14 technique is now standardized in a host of international standards including ASTM D6866, CEN 16137, EN 16640, ISO 16620, ISO 19984, BS EN ISO 21644:2021, ISO 13833 and EN 16785. Carbon-14 analysis can be used on various types of samples (gas, liquids and solids). Beta Analytic continues to be a technical contact for ASTM D6866 with current president Ron Hatfield and is involved with all their latest ASTM D6866 versions.

The Carbon-14 standardized method is also incorporated in a variety of regulatory programs including the California AB32 program, US EPA GHG Protocol, US EPA Renewable Fuels Standard, United Nations Carbon Development Mechanism, Western Climate Initiative, Climate Registry's Greenhouse Gas Reporting Protocol and EU Emissions Trading Scheme.

We are currently technical experts on Carbon-14 in the following committees:

ASTM D6866 (D20.96) Plastics and Biobased Products (Technical Advisor)
ASTM (D02.04) Petroleum Products, Liquid Fuels and Lubricants (Technical Advisor)
ASTM (061) US TAG to ISO/TC 61 Plastics (Technical Expert)
USDA BioPreferred Program TAC (Technical Advisor)
ISO/TC 61/SC14/WG1 Terminology, classifications, and general guidance (Technical Expert)
CEN/TC 411 Biobased Products
CEN/TC 411/WG 3 Biobased content
CEN/TC 61/SC 14/WG 1 Terminology, classifications, and general guidance (Technical Expert)

²⁷2022. "Testing the methods for determination of radiocarbon content in liquid fuels in the Gliwice Radiocarbon and Mass Spectrometry Laboratory." *Radiocarbon*



ISO/IEC 17025:2017-Accredited Testing Laboratory

ISO/IEC 17025:2017 Accredited Laboratory

To ensure the highest level of quality, laboratories performing ASTM D6866 testing should be ISO/IEC 17025:2017 accredited or higher. This accreditation is unbiased, third party awarded and supervised. It is unique to laboratories that not only have a quality management program conformant to the ISO 9001:2008 standard, but more importantly, have demonstrated to an outside third-party laboratory accreditation body that Beta Analytic has the technical competency necessary to consistently deliver technically valid test results. The ISO 17025 accreditation is specifically for natural level radiocarbon activity measurements including biobased analysis of consumer products and fuels, and for radiocarbon dating.

Required tracer-free facility for Carbon-14

For carbon-14 measurement to work, be accurate, and repeatable, the facility needs to be a tracer-free facility, which means artificial/labeled carbon-14 is not and has never been handled in that lab. Facilities that handle artificial carbon-14 use enormous levels relative to natural levels and it becomes ubiquitous in the facility and cross contamination within the facility, equipment and chemistry lines is unavoidable. Results from a facility that handles artificial carbon-14 would show elevated renewable contents (higher pMC, % Biobased / Biogenic values), making those results invalid. Because of this, Federal contracts and agency programs (such as the USDA BioPreferred Program) require that AMS laboratories must be 14C tracer-free facilities in order to be considered for participation in solicitations.

Areas where cross-contamination might occur include but are not limited to; biomedical or nuclear reactors, isotope enrichment / depletion columns, water, soil, plant, or air samples collected near or at biomedical / nuclear reactor sites, medical, industrial, or hazardous waste sites, samples specifically manipulated to study the uptake / fractionation of stable isotopes due to biological or metabolic processes. To learn more about the risks associated with testing natural levels Carbon-14 samples in a facility handling artificially enhanced isotopes please see the additional information provided after this comment.

References

2006. "Determining the modern carbon content of biobased products using radiocarbon analysis." *Bioresource Technology*, 97(16), 2084-2090.
2010. "40 CFR Part 80 Subpart M— Renewable Fuel Standard." *National Archives Code of Federal Regulations* <https://www.ecfr.gov/current/title-40/chapter-I/subchapter-C/part-80/subpart-M>
2013. "ISO 13833:2013 Stationary source emissions: Determination of the ratio of biomass (biogenic) and fossil-derived carbon dioxide." *International Organization for Standardization*



ISO/IEC 17025:2017-Accredited Testing Laboratory

2020. "Reporting Co-Processing and Renewable Gasoline Emissions Under MRR." *California Air Resources Board*
https://ww2.arb.ca.gov/sites/default/files/2020-09/MRR_coprocessing-slides_Sept_2020.pdf
2021. "Standard Test Methods for Determining the Biobased Content of Solid, Liquid, and Gaseous Samples Using Radiocarbon Analysis." *ASTM International (D6866-21)*. pp 1-19. doi: 10.1520/D6866-21.
2022. "Clean Fuel Regulations: Quantification Method for Co-Processing in Refineries." *Environment and Climate Change Canada*
<https://www.canada.ca/en/environment-climate-change/services/managing-pollution/energy-production/fuel-regulations/clean-fuel-regulations/compliance/quantification-methodco-processing-refineries.html>
2022. "Testing the methods for determination of radiocarbon content in liquid fuels in the Gliwice Radiocarbon and Mass Spectrometry Laboratory." *Radiocarbon*, 64(6), pp.1-10. DOI:10.1017/RDC.2022.35
2022. "Chapter 173-424 WAC: Clean Fuels Program Rule." *Washington State Legislature*
<https://app.leg.wa.gov/WAC/default.aspx?cite=173-424&full=true>
2023. "Renewable energy- method for calculating the share of renewables in the case of co-processing." *European Commission*
https://ec.europa.eu/info/law/better-regulation/have-your-say/initiatives/12711-Renewable-energy-method-for-calculating-the-share-of-renewables-in-the-case-of-co-processing_en
2023. "40 CFR Parts 80 and 1090– Renewable Fuel Standard (RFS) Program: Standards for 2023–2025 and Other Changes." *Environmental Protection Agency*
<https://www.govinfo.gov/content/pkg/FR-2023-07-12/pdf/2023-13462.pdf>
2023. "Oregon Clean Fuels Program." *Oregon Department of Environmental Quality*
<https://secure.sos.state.or.us/oard/displayDivisionRules.action?selectedDivision=1560>
2023. "RSB Standard for Advanced Fuels." *Roundtable on Sustainable Biomaterials (RSB)*
https://rsb.org/wp-content/uploads/2024/03/RSB-STD-01-010-RSB-Standard-for-advanced-fuels_v2.6-1.pdf
2023. "RR:D02-2052." *ASTM International* <https://www.astm.org/rr-d02-2052.html>
2023. "ISCC Press Release July 27, 2023." *International Sustainability & Carbon Certification*
<https://www.iscc-system.org/news/press-release-27-july-2023/>
2024. "Biden EPA Rejects Plastics Industry's Fuzzy Math That Misleads Customers About Recycled Content." *ProPublica* <https://www.propublica.org/article/epa-rejects-mass-balance-plastics-recycling-safer-choice>
2024. "ExxonMobil Accused of "Deceptively" Promoting Chemical Recycling as a Solution for the Plastics Crisis." *ProPublica* <https://www.propublica.org/article/exxonmobil-plastics-recycling-pyrolysis-lawsuit-california>
2024. "The People of the State of California v. Exxon Mobil Corporation." *Superior Court of the State of California*



https://oag.ca.gov/system/files/attachments/press-docs/Complaint_People%20v.%20Exxon%20Mobil%20et%20al.pdf

2024. "Is Your Water Bottle Really Made From Recycled Plastic?" *The New York Times*

https://www.nytimes.com/2024/08/26/business/energy-environment/tritan-renew-plastic-bottles-recycled.html?unlocked_article_code=1.F04.bY6M.-T BnLuNxj6i&smid=url-share

2024. "The Mass Balance Approach." *International Sustainability & Carbon Certification*

<https://www.iscc-system.org/news/mass-balance-explained/#:~:text=Mass%20balance%20provides%20manufacturers%20with,production%20process%20through%20certified%20bookkeeping.>

2024. "Opinion of Advocate General Campos Sánchez-Bordona Delivered on 11 January 2024: Case C-624/22." *Court of Justice of the European Union*

<https://eur-lex.europa.eu/legal-content/EN/TXT/HTML/?uri=CELEX:62022CC0624#Footnote1>

2024. "Board Members of the ISCC Association." *International Sustainability & Carbon Certification*

<https://www.iscc-system.org/governance/iscc-association/board/>

2025. "Low Carbon Fuel Regulation: Co-Processing Methodology" *British Columbia Ministry of Energy and Climate Solutions*

https://www2.gov.bc.ca/assets/gov/farming-natural-resources-and-industry/electricity-alternative-energy/transportation/renewable-low-carbon-fuels/rlcf019_-_coprocessing_methodology_final.pdf

Demand a Tracer-Free Laboratory for Radiocarbon Dating

As part of its commitment to provide high-quality results to its clients, ISO/IEC 17025-accredited Beta Analytic does not accept pharmaceutical samples with “tracer Carbon-14” or any other material containing artificial Carbon-14 (^{14}C) to eliminate the risk of cross-contamination. Moreover, the lab does not engage in “satellite dating” – the practice of preparing individual sample graphite in a remote chemistry lab and then subcontracting an AMS facility for the result.

High Risk of Cross-Contamination

Pharmaceutical companies evaluate drug metabolism by using a radiolabeled version of the drug under investigation. AMS biomedical laboratories use ^{14}C as a tracer because it can easily substitute ^{12}C atoms in the drug molecule, and it is relatively safe to handle. Tracer ^{14}C is a well-known transmittable contaminant to radiocarbon samples, both within the AMS equipment and within the chemistry lab.

Since the artificial ^{14}C used in these studies is phenomenally high (enormous) relative to natural levels, once used in an AMS laboratory it becomes ubiquitous. Cross-contamination within the AMS and the chemistry lines cannot be avoided. Although the levels of contamination are acceptable in a biomedical AMS facility, it is not acceptable in a radiocarbon dating facility.

Biomedical AMS facilities routinely measure tracer-level, labeled (Hot) ^{14}C samples that are hundreds to tens of thousands of times above the natural ^{14}C levels found in archaeological, geological, and hydrological samples. Because the ^{14}C content from the biomedical samples is so high, even sharing personnel will pose a contamination risk; “Persons from hot labs should not enter the natural labs and vice versa” (Zermeño et al. 2004, pg. 294). These two operations should be absolutely separate. Sharing personnel, machines, or chemistry lines run the risk of contaminating natural level ^{14}C archaeological, geological, and hydrological samples.

Avoid the Risks

Find out from the lab that you are planning to use that they have never in the past and will never in the future:

- accept, handle, graphitize or AMS count samples containing Tracer or Labeled (Hot) ^{14}C .

- share any laboratory space, equipment, or personnel with anyone preparing (pretreating, combusting, acidifying, or graphitizing) samples that contain Tracer or Labeled (Hot) ^{14}C .

- use AMS Counting Systems (including any and all beam-line components) for the measurement of samples that contain Tracer or Labeled (Hot) ^{14}C .

Tracer-Free Lab Required

Recently, federal contracts are beginning to specify that AMS laboratories must be ^{14}C tracer-free facilities in order to be considered for participation in solicitations.

A solicitation for the National Oceanic and Atmospheric Administration (NOAA) has indicated that “the AMS Facility utilized by the Contractor for the analysis of the micro-samples specified must be a ^{14}C tracer-level-free facility.” (Solicitation Number: WE-133F-14-RQ-0827 - Agency: Department of Commerce)

As a natural level radiocarbon laboratory, we highly recommend that researchers require the AMS lab processing their samples to be Tracer-free.

No Exposure to Artificial Carbon-14

According to ASTM International, the ASTM D6866 standard is applicable to laboratories working without exposure to artificial carbon-14 routinely used in biomedical studies. Artificial carbon-14 can exist within the laboratory at levels 1,000 times or more than 100 % biobased materials and 100,000 times more than 1% biobased materials. Once in the laboratory, artificial ^{14}C can become undetectably ubiquitous on materials and other surfaces but which may randomly contaminate an unknown sample producing inaccurately high biobased results. Despite vigorous attempts to clean up contaminating artificial ^{14}C from a laboratory, isolation has proven to be the only successful method of avoidance. Completely separate chemical laboratories and extreme measures for detection validation are required from laboratories exposed to artificial ^{14}C . Accepted requirements are:

- (1) disclosure to clients that the laboratory working with their products and materials also works with artificial ^{14}C
- (2) chemical laboratories in separate buildings for the handling of artificial ^{14}C and biobased samples
- (3) separate personnel who do not enter the buildings of the other
- (4) no sharing of common areas such as lunch rooms and offices
- (5) no sharing of supplies or chemicals between the two
- (6) quasi-simultaneous quality assurance measurements within the detector validating the absence of contamination within the detector itself.

ASTM D6866-22 – Standard Test Methods for Determining the Biobased Content of Solid, Liquid, and Gaseous Samples Using Radiocarbon Analysis.

Useful Reference

1. Memory effects in an AMS system: Catastrophe and Recovery. J. S. Vogel, J.R. Southon, D.E. Nelson. Radiocarbon, Vol 32, No. 1, 1990, p. 81-83 doi:10.2458/azu_js_rc.32.1252 (Open Access)

"... we certainly do not advocate processing both labeled and natural samples in the same chemical laboratory." "The long term consequences are likely to be disastrous."

2. Recovery from tracer contamination in AMS sample preparation. A. J. T. Jull, D. J. Donahue, L. J. Toolin. Radiocarbon, Vol. 32, No.1, 1990, p. 84-85 doi:10.2458/azu_js_rc.32.1253 (Open Access)

"... tracer ^{14}C should not be allowed in a radiocarbon laboratory." "Despite vigorous recent efforts to clean up the room, the "blanks" we measured had ^{14}C contents equivalent to modern or even post -bomb levels."

3. Prevention and removal of elevated radiocarbon contamination in the LLNL/CAMS natural radiocarbon sample preparation laboratory. Zerneño, et. al. Nuclear Instruments and Methods in Physics Research Section B: Beam Interactions with Materials and Atoms Vol. 223-224, 2004, p. 293-297 doi: 10.1016/j.nimb.2004.04.058

"The presence of elevated ^{14}C contamination in a laboratory preparing samples for natural radiocarbon analysis is detrimental to the laboratory workspace as well as the research being conducted."

4. High level ^{14}C contamination and recovery at XI'AN AMS center. Zhou, et. al. Radiocarbon, Vol 54, No. 2, 2012, p. 187-193 doi:10.2458/azu_js_rc.54.16045

"Samples that contain high concentrations of radiocarbon ("hot" samples) are a catastrophe for low background AMS laboratories." "In our case the ion source system was seriously contaminated, as were the preparation lines."



Beta Analytic

www.radiocarbon.com